

**Test Parameters**

Control Number: essai
 Type of Test: Flow
 Date Tested: 09-Apr-2020
 Start Time: 16:54:32
 Tested by: PESTEL

Test Instrument

Fluke Biomedical IDA-5
 Infusion Device Analyzer
 Serial Number: 4603553
 Version: Ver 1.37
 Channel 1 S/N 07190099 Ver 1.05

Pump Details

Manufacturer: ELECTROLAB
 Device Type (Model): Prototype 1
 Serial Number:
 Location:
 Comments:

Set Values

Flow Rate: 3.00 ml/h
 Total Volume: 40.00 ml

Measured Values

Flow Rate: 3.01 ml/h
 Total Volume: 38.03 ml
 Total Test Time: 12:39:01