

**Test Parameters**

Control Number ESSAI
 Type of Test Flow
 Date Tested 09-Apr-2020
 Start Time 11:21:54
 Tested by PESTEL

Test Instrument

Fluke Biomedical IDA-5
 Infusion Device Analyzer
 Serial Number 4603553
 Version Ver 1.37
 Channel 1 S/N 07190099 Ver 1.05

Pump Details

Manufacturer ELECTROLAB
 Device Type (Model) Prototype 1
 Serial Number
 Location
 Comments

Set Values

Flow Rate 50.00 ml/h
 Total Volume 50.00 ml

Measured Values

Flow Rate 50.16 ml/h
 Total Volume 41.99 ml
 Total Test Time 00:50:13