

**Test Parameters**

Control Number	essai
Type of Test	Occlusion
Date Tested	10-Apr-2020
Start Time	12:04:24
Tested by	PESTEL

Test Instrument

Fluke Biomedical IDA-5	
Infusion Device Analyzer	
Serial Number	4603553
Version	Ver 1.37
Channel 1	S/N 07190099
	Ver 1.05

Pump Details

Manufacturer	ELECTROLAB
Device Type (Model)	Prototype 1
Serial Number	
Location	
Comments	

Set Values

Flow Rate	50.00 ml/h
-----------	------------

Measured Values

Peak Pressure	487 mmHg
Peak Pressure	9.42 psi
Peak Pressure	64.94 kPa
Occurred After	00:02:20
Total Test Time	00:04:09