

**Test Parameters**

Control Number essai
Type of Test Occlusion
Date Tested 10-Apr-2020
Start Time 12:12:46
Tested by PESTEL

Test Instrument

Fluke Biomedical IDA-5
Infusion Device Analyzer
Serial Number 4603553
Version Ver 1.37
Channel 1 S/N 07190099 Ver 1.05

Pump Details

Manufacturer ELECTROLAB
Device Type (Model) Prototype 1
Serial Number
Location
Comments

Set Values

Flow Rate 5.00 ml/h

Measured Values

Peak Pressure 601 mmHg
Peak Pressure 11.62 psi
Peak Pressure 80.14 kPa
Occurred After 00:34:55
Total Test Time 00:44:04